

Technical data

ProSim 8 / ProSim 8P Vital Signs and ECG Patient Simulator



Leading Edge Performance in Vital Signs Simulation

The industry-leading ProSim 8 / ProSim 8P patient simulator enables biomedical technicians to complete preventive maintenance testing with one device in less than five minutes. The ProSim 8 / ProSim 8P patient simulator is ideal for biomedical technicians testing patient monitors for certain known physiological output signals.

Key features

- 8-in-1 multifunction patient simulator is a combination ECG simulator, fetal simulator, arrhythmia simulator, respiration simulator, temperature simulator, IBP simulator, cardiac output simulator, cardiac catheterization simulator, NIBP simulator
- SpO₂ simulator to test Rainbow multi-wavelength waveforms (with optional ProSim SpO₂ Test Module)
- Stay-connected ECG posts for easy/secure ECG snap and lead connections
- Static pressure linearity testing
- Improved manometer accuracy with the ProSim 8P
- Repeatable NIBP simulation for dynamic pressure repeatability testing
- Physiologically synchronized pulses across all parameters
- Barcode scanning and direct data capture and printing functionality
- Onboard, patient pre-sets and auto sequences for fast/easy testing
- Multi-language user interface offers choice of language selection
- The device includes an external, long-life battery that allows for more flexible charging options
- Optional OneQA Workflow Automation Software offers customizable procedures/checklists and automated data capture/storage

Specifications

General specifications		
Dimensions (L x W x H)	14.5 cm x 30.2 cm x 8.6 cm (5.7 in x 11.9 in x 3.4 in)	
Display	LCD color display	
Communication	USB device upstream port	Mini-B connector for control by a computer
	USB host controller port	Type A, 5 V output, 0.5 A max load. Connector for keyboard, barcode reader, and printer
Battery life	9 hours (minimum), 100 NIBP cycles typical	
Weight	1.81 kg (4 lb)	
Safety standards	IEC/EN 61010-1 3rd Edition; Pollution degree 2 CAT None	
Certifications	CE, CSA, C-TICK N10140, RoHS	
Electromagnetic compatibility (EMC)	IEC 61326-1:2006	
Detailed specifications		
Normal-sinus-rhythm waveform		
ECG reference	The ECG amplitudes specified are for Lead II (calibration), from the baseline to the peak of the R wave. All other leads are proportional	
Normal sinus rhythm	12-lead configuration with independent outputs referenced to right leg (RL). Output to 10 universal ECG jacks, color-coded to AHA and IEC standards	
High-level output	0.5 V/mV ± 5 % of the ECG amplitude setting available on a BNC connector	
Amplitude	0.05 mV to 0.5 mV (0.05 mV steps); 0.5 mV to 5.0 mV (0.25 mV steps) Other leads are proportional to Lead II (reference lead) in percentage per:	
	Lead I: 70 Lead II: 100 Lead III: 30	Lead V1: 24 Lead V2: 48 Lead V3: 100 Lead V4: 120 Lead V5: 112 Lead V6: 80
Amplitude accuracy	ProSim 8	± (2 % of setting + 0.05 mV)
	ProSim 8P	±2 % of setting (0.5 mV to 5.0 mV, Lead II) ±(2 % of setting + 0.05 mV) (0.05 mV to 0.5 mV, Lead II; 0.05 mV to 5.0 mV, all other leads)
ECG rate	10 BPM to 360 BPM in 1 BPM steps	
Rate accuracy	± 1 % of setting	
ECG waveform selection	Adult (80 ms) or pediatric (40 ms) QRS duration	
ST-segment elevation	Adult mode only. -0.8 mV to +0.8 mV (0.1 mV steps). Additional steps: + 0.05 mV and - 0.05 mV Based on Lead II 1.0mV nominal waveform, proportional to selected amplitude.	
Power-on default	60 BPM, 1.0 mV, adult QRS and ST-segment elevation of 0 mV	



Pacemaker waveform		
Pacer pulse	Amplitude	0 (off), ± 2, ± 4, ± 6, ± 8, ± 10, ± 12, ± 14, ± 16, ± 18, ± 20, ± 50, ± 100, ± 200, ± 500, and ± 700 mV for lead II (reference lead)
	Accuracy	Reference lead II: ± (5 % setting + 0.2 mV)
		All other leads: ± (10 % setting + 0.4 mV)
Pacer pulse width	0.1 ms, 0.2 ms, 0.5 ms, 1 ms, and 2 ms ± 5 %	
Paced arrhythmias	Atrial 80 BPM Asynchronous 75 BPM Demand with frequent sinus beats Demand with occasional sinus beats Atrio-ventricular sequential Noncapture (one time) Nonfunction	
Power-on default	Amplitude 5 mV, width 1 ms, atrial waveform	
Arrhythmia		
Baseline NSR	80 BPM	
PVC focus	Left focus, standard timing (except where specified)	
Supraventricular arrhythmia	Atrial fibrillation (coarse or fine); atrial flutter; sinus arrhythmia; missed beat (one time); atrial tachycardia; paroxysmal atrial tachcardia; nodal rhythm; and supraventricular tachycardia	
Premature arrhythmia	Premature atrial contraction (PAC); premature nodal contraction (PNC); PVC1 left ventricular; PVC1 left ventricular, early; PVC1 left ventricular, R on T; PVC2 right ventricular; PVC2 right ventricular, early; PVC2 right ventricular, R on T; and multifocal PVCs	
Ventricular arrhythmia	PVCs 6, 12, or 24 per minute; frequent multifocal PVCs; bigeminy; trigeminy; multiple PVCs (one-time run of 2, 5, or 11 PVCs); monoventricular tachycardia (120 to 300 BPM in 5 BPM steps); poly-ventricular tachycardia (5 types); ventricular fibrillation (coarse or fine); and asystole	
Conduction defect	First-, second-, or third-degree heart block; and right- or left-bundle branch block	
Advanced cardiac life support	Shockable pulseless arrest rhythms	Ventricular fibrillation (coarse), ventricular fibrillation (fine), unstable polymorphic ventricular tachycardia
	Non-shockable pulseless arrest rhythms	Asystole
	Symptomatic bradycardia	Sinus bradycardia (< 60 BPM) 2nd degree AV block, mobitz type I 2nd degree AV block, mobitz type II Complete/3rd degree AV block Right bundle branch block Left bundle branch block
		Sinus tachycardia > 150 BPM
		Supraventricular Tachycardia
	Symptomatic tachycardia: regular narrow-complex tachycardia (QRS < 0.12 seconds)	Sinus tachycardia > 150 BPM
		Supraventricular tachycardia SVT with aberrancy
	Symptomatic tachycardia: regular wide-complex tachycardias (QRS ≥ 0.12 seconds)	Sinus tachycardia > 150 BPM
	Irregular tachycardia	Atrial fibrillation (coarse and fine), atrial flutter, unstable monomorphic ventricular tachycardia (120 BPM to 300 BPM), torsade de pointes/polymorphic ventricular tachycardia (long QT interval)

ECG performance testing

Amplitude	0.05 mV to 0.5 mV (0.05 mV steps); 0.5 mV to 5.0 mV (0.25 mV steps) Other leads are proportional to Lead II (reference lead) in percentage per: Lead I: 70 Lead III: 30 Lead II: 100 Lead V1 through V6: 100	
Pulse wave	30 BPM, 60 BPM, with 60 ms pulse width	
Square wave	0.125 Hz, 2 Hz, 2.5 Hz	
Triangle wave	0.125 Hz, 2 Hz, 2.5 Hz	
Sine wave	0.05 Hz, 0.5 Hz, 1, 2 Hz, 5 Hz, 10 Hz, 25 Hz, 30 Hz, 40 Hz, 50 Hz, 60 Hz, 100 Hz, and 150 Hz	
R-wave detection	Waveform	Triangular pulse
	Rate	30 BPM, 60 BPM, 80 BPM, 120 BPM, 200 BPM, and 250 BPM
	Width	8 ms to 20 ms in 2 ms steps, and 20 ms to 200 ms in 10 ms steps
	Width accuracy	± (1 % of setting + 1 ms)
QRS detection	Widths	8 ms to 20 ms in 2 ms steps, and 20 ms to 200 ms in 10 ms steps
	Width accuracy	± (1 % of setting + 1 ms)
	Rate	30 BPM, 60 BPM, 80 BPM, 120 BPM, 200 BPM, and 250 BPM
	R-Wave up slope	0.875 amplitude, 0.4375 x width
	R-Wave down slope	Full amplitude, 0.5 x width
	S-Wave up slope	0.125 amplitude, 0.0625 x width
Tall T-wave rejection	Waveform	QT Interval 350 ms T-Wave width 180 ms T-Wave shape ½ sinewave
	Amplitude	0 % to 150 % reference lead amplitude in 10 % steps
	Rate	80 BPM
Rate accuracy	± 1 % of setting	
Amplitude accuracy	ProSim 8	± (2 % of setting + 0.05 mV)
	ProSim 8P	±2 % of setting (0.5 mV to 5.0 mV, Lead II) ±(2 % of setting + 0.05 mV) (0.05 mV to 0.5 mV, Lead II; 0.05 mV to 5.0 mV, all other leads)

ECG artifact

Type	50 Hz, 60 Hz, muscular, baseline wander, respiration
Size	25 %, 50 %, 100 % of the normal sinus R-Wave for each lead
Lead select	All, RA, LL, LA, V1, V2, V3, V4, V5, V6

Fetal/Maternal ECG

Fetal heart rate (fixed)	60 BPM to 240 BPM in 1 BPM steps
Fetal heart rate (IUP)	140 BPM at beginning, then varies with pressure
Intrauterine-pressure waveforms	Early deceleration, late deceleration, and acceleration
Wave duration	90 seconds, bell-shaped pressure curve, from 0 mmHg to 90 mmHg and returning to 0
IUP period	2 min, 3 min, or 5 minutes; and manual
Default settings	FHR 140 BPM, early deceleration wave, manual

Invasive blood pressure		
Channels	2, each independently settable with identical parameters and are individually electrically isolated from all other signals	
Input/output impedance	300 Ω ± 10 %	
Exciter input range	2 to 16 V peak	
Exciter-input frequency range	DC to 5000 Hz	
Transducer sensitivity	5 (default) or 40 μV/V/mmHg	
Pressure accuracy	± (1 % of setting + 1 mmHg) accuracy guaranteed for dc excitation only	
Static pressure	- 10 to + 300 mmHg in 1 mmHg steps	
Pressure units	mmHg or Kpa	
Dynamic waveforms	Types (default pressures	Arterial (120/80)
		Radial artery (120/80)
		Left ventricle (120/00)
		Right ventricle (25/00)
		Pulmonary artery (25/10)
		Pulmonary-artery wedge (10/2)
		Right atrium (central venous or CVP) (15/10)
	Pressure variability	Systolic and diastolic pressures are independently variable in 1 mmHg steps
Swan-Ganz sequence	Right atrium, right ventricle (RV), pulmonary artery (PA), pulmonary artery wedge (PAW)	
Cardiac catheterization	Chambers	Aortic, pulmonary valve, and mitral valve
Respiration artifact	Arterial, radial artery, and left ventricle	5 % to 10 % multiplication
BP output	Circular DIN 5-Pin	
Power-on default	0 mmHg	
Respiration		
Rate	0 (OFF), 10 BrPM to 150 BrPM in 1 BrPM steps	
Waves	Normal or ventilated	
Ratio (inspiration:expiration)	Normal	1:1, 1:2, 1:3, 1:4, 1:5
	Ventilated	1:1
Impedance variations (Δ Ω)	0.00 Ω to 1.00 Ω in 0.05 Ω steps and 1 Ω to 5 Ω in 0.25 Ω steps	
Accuracy delta	± (5 % of setting + 0.1 Ω)	
Baseline	500 Ω, 1000 Ω (default), 1500 Ω, 2000 Ω, Leads I, II, III	
Accuracy baseline	± 5 %	
Respiration lead	LA or LL (default)	
Apnea selection	12 sec, 22 sec, or 32 seconds (one-time events), or continuous (Apnea ON = respiration OFF)	
Power-on default	20 BrPM, delta 1.0 Ω	
Temperature		
Temperature	30.0 °C to 42.0 °C in 0.5 °C steps for standard temperature. 0 °C, 24 °C, 25 °C, 37 °C, and 40 °C for high accuracy temperature. ¹	
Accuracy	±0.4 °C for standard temperature. ±0.1 C for high accuracy temperature ¹	
Compatibility	Yellow Springs, Inc. (YSI) Series 400 and 700	
Output	Circular DIN 4-Pin	

1. Units made before 2025 may not have the High Accuracy mode. Check the product to determine if it does.

Cardiac output

Catheter type	Baxter Edwards, 93a-131-7f
Calibration coefficient	0.542 (0 °C injectate), 0.595 (24 °C injectate)
Blood temperature	36 °C (98.6 °F) to 38 °C (100.4 °F) ± 0.2 °C in 1 °C steps
Injectate volume	10 cc
Injectate temperature	0 °C or 24 °C
Cardiac output	2.5, 5, 10 liters per minute ± 7.5 %
Faulty-injectate curve	Waveform for simulation available
Left-to-right-shunt curve	Waveform for simulation available
Calibrated pulse	1.5 ° for 1 second
Connector	Circular DIN 7 pin
Power-on default	5 liters per minute, 0 °C injectate, 37 °C blood temperature

Non-invasive blood pressure

Pressure units	mmHg or kPa	
Manometer (pressure meter)	ProSim 8 Range	10 mmHg to 400 mmHg
	ProSim 8 Accuracy	±(0.5 % reading + 0.5 mmHg)
	ProSim 8P Range	0 mmHg to 400 mmHg
	ProSim 8P Accuracy	0 mmHg to 300 mmHg (±0.75 mmHg) ¹ 300 mmHg to 400 mmHg ±(0.5 % reading + 0.5 mmHg)
	Resolution (ProSim 8 and 8P)	0.1 mmHg
Pressure source	Target pressure range	15 mmHg to 400 mmHg
	Resolution	1 mmHg
NIBP simulations	Pulse	2 mmHg max into 500 ml NIBP system
	Volume of air moved	1.25 ml max
	Simulations (systolic/diastolic [MAP])	Adult: 60/30 (40), 80/50 (60); 100/65 (77); 120/80 (93); 150/100 (117); and 200/150 (167) and 255/195 (215) Neonatal: 35/15 (22); 60/30 (40); 80/50 (60); 100/65 (77); 120/80 (93) and 150/100 Pressure variability: systolic and diastolic pressures are variable by 1 mmHg
	Repeatability	Within ± 2 mmHg (at maximum pulse size independent of device under test)
	Synchronization: normal Sinus heart rates: 30 BPM to 240 BPM	Maximum rate at 1 ml: 240 BPM achievable with pulses up to 1 ml Maximum rate at 1.25 ml: 180 BPM
	Synchronization: arrhythmias	Premature atrial contraction (PAC), premature ventricular contraction (PVC), atrial fibrillation, and missed beat
Leak test	Target pressure	15 mmHg to 400 mmHg
	Elapse time	0:30 min to 5:00 minutes: seconds in 30 second steps
	Leakage rate	0 mmHg/minute to 200 mmHg/minute <2 mmHG/min into 500 ml rigid volume
Pressure relief test range	100 to 400 mmHg	

1. To achieve this accuracy, the product must be operating between 10 °C and 30 °C (50 °F and 86 °F) and the user must zero the manometer manually before each test.

Oximeter SpO₂ optical emitter and detector (optional)

% O ₂	Range	30 % to 100 %
	Resolution	1 %
% O ₂ accuracy	With oximeter manufacturer's R-curve	Saturation within UUT specific range: ± (1 count + specified accuracy of the UUT) Saturation outside UUT specific range: monotonic with unspecified accuracy
	With Fluke Biomedical R-curves	91 % to 100 % ± (3 counts + specified accuracy of the UUT)
		81 % to 90 % ± (5 counts + specified accuracy of the UUT)
		71 % to 80 % ± (7 counts + specified accuracy of the UUT)
		Below 71 % monotonic with unspecified accuracy
% O ₂ repeatability (ProSim 8P) ²	≤1% (Calculated as 1 standard deviation of repeated measurements, independent of device under test)	
Heart rate	30 BPM to 300 BPM in 1 BPM steps. Oximeter SpO ₂ optical emitter and detector is synchronized with ECG rate delayed by 150 ms.	
Transmission: ratio of detector current to LED current, expressed in parts per million (ppm)	Range	0 ppm to 300.00 ppm
	Resolution	0.01 ppm
	Accuracy	+ 50 %/- 30 % for compatible monitors, unspecified for others. Selected by finger size and color: dark, thick finger, medium finger, light, thin finger, neonatal foot.
Pulse amplitude	Range	0 % to 20.00 %
	Resolution	0.01 %
Artifact	Respiration	Range: 0 % to 5 % of transmission
		Resolution: 1 %
		Rate: all ProSim respiration simulation settings
	Ambient light	Range: 0 to 5X transmitted light
		Resolution: 1X Frequency: DC, 50 Hz, 60 Hz, and 1 kHz to 10 kHz in 1 kHz steps
Masimo Rainbow technology	Masimo Rainbow technology with an optional adapter cable supplied by Masimo that allows the ProSim two wavelength to test the Rainbow multiple wavelength system	
Compatible manufacturer products	With manufacturer R-curve	Nellcor, Masimo, Nonin, and Nihon Kohden
	With Fluke R-curve	Mindray, GE-Ohmeda, Philips/HP, and BCI

2. This specification only applies to the ProSim 8P



Pre-Defined Simulations

Normal
Hypertensive
Hypotensive
Tachycardic
Bradycardic
Ventricular fibrillation
Asystole

Autosequences (default)

Monitor testing sequence
ECG testing
Oximeter testing sequence
Cardiac failure sequence
Arrhythmia sequence
Exercise sequence
Respiration sequence
Performance wave test
IBP testing sequence
Temperature sequence

Ordering information

Part number	Description
3985658	ProSim SpO ₂ Test Module
4034609	ProSim Rainbow Test Cable
3979409	ProSim 8 Vital Signs Simulator
3985195	ProSim 8 Vital Signs Simulator, AUSTRALIA
3985206	ProSim 8 Vital Signs Simulator, BRAZIL
3985188	ProSim 8 Vital Signs Simulator, JAPAN
3985163	ProSim 8 Vital Signs Simulator, SCHUKO
3985174	ProSim 8 Vital Signs Simulator, UK
5255933	ProSim 8 Vital Signs Simulator, AUS + SPOT
5255940	ProSim 8 Vital Signs Simulator, BRAZ + SPOT
5262232	ProSim 8 Vital Signs Simulator, JAPAN + SPOT
5255916	ProSim 8 Vital Signs Simulator, SCHUKO + SPOT
5255925	ProSim 8 Vital Signs Simulator, UK + SPOT
6021706	ProSim 8P Vital Signs Simulator
4322373	ProSim 8P Vital Signs Simulator, AUSTRALIA
6021982	ProSim 8P Vital Signs Simulator, BRAZIL
6021995	ProSim 8P Vital Signs Simulator, JAPAN
4322364	ProSim 8P Vital Signs Simulator, SCHUKO
6022079	ProSim 8P Vital Signs Simulator, UK
6024669	ProSim 8P Vital Signs Simulator + SPOT
6022257	ProSim 8P Vital Signs Simulator, AUS + SPOT
6022274	ProSim 8P Vital Signs Simulator, BRAZ + SPOT
6022288	ProSim 8P Vital Signs Simulator, JAPAN + SPOT
6022290	ProSim 8P Vital Signs Simulator, SCHUKO + SPOT
6022300	ProSim 8P Vital Signs Simulator, UK + SPOT

Standard Accessories

Part number	Description
3980671	ProSim 8 / 8P Users Manual
3980667	ProSim 8 / 8P Getting start manual
4021085	ProSim 8 / 8P Battery Pack
4034393	USB Cable
2392173	IBP Cable, unterminated
4034597	ProSim 8 / 8P Carrying Case
4308086	ProSim NIBP Mandrel Set
2391882	Set of NIBP Cuff Adapters
2184298	AC/DC Power Supply Power cord (country-specific)



Optional Accessories

Part number	Description
2392199	CI-3 Cardiac Output Box
3408564	Mini-DIN to DIN IBP Adapter
4034611	NIBP Rigid Test Chamber 500ML
4034627	Ansur Test Software ProSim 8 Plug-In

Cable kits

Part number	Model	Includes
3984910	ProSim 8 Accessory Kit	DIN to minDin adapter, HP/Philips Intellivue IBP cable, GE Marquette Eagle/Dash/Solar IBP cable, Welch Allyn Propaq/SpaceLabs Ultraview IBP cable, YSI400 series temperature cable, YSI700 series temperature cable, CI-3 Cardiac Output Box, spare battery pack)
3984922	HP/Phillips intellivue Cable Set	HP-3 BP Cable (2198902) two, HPT-2 Tamp/C.O. Injct Cable Assembly (2199257), COA-1 Cable Assembly (2199240), UT-4, Low profile 1/4 inch phone plug, YSI 400 Series Compatible 2 conductor (2523334)
3984968	GE Marquette Eagle/Dash/Solar Cable Set	MQ-3 BP Cable (2199627) two, UT-4 Low profile 1/4 inch phone plug, YSI 400 series Compatible 2 conductor (2523334), UT-2 Tamp Cable 700 series YSI (2199019), PROSIM8- 4402GECO, Din cardiac Output Marq Eagle (4022300)
3984946	ProSim 8 SpaceLabs Ultraview Cable Set	TK-1 BP Cable (2198879) two, UT-4 Low profile 1/4 inch phone plug, YSI 400 series compatible 2 conductor (2523334), UT-2 Tamp Cable 700 Series YSI (2199019)
3984979	Welch Allyn/Propaq Cable Set	TK-1 BP Cable (2198879) two, UT-4 Low profile 1/4 inch phone plug, YSI 400 Series Compatible 2 conductor (2523334), UT-2 Tamp cable 700 series YSI (2199019)
3984993	Drager Infinity Cable Set	SM-1 BP Cable (2198925) two, UT-4, Low profile 1/4 inch phone plug, YSI 400 series compatible 2 conductor (2523334)
3985009	ProSim 8 Nihon Kohden Cable Set	Nihon Kohden-NK-1, BP Cable (5M) (2462263) two, DIN to mindin adapter (3408564), UT-4, Low profile 1/4 inch phone plug, YSI 400 Series Compatible 2 conductor (2523334)

The ProSim 8 does not provide simulation for all types of fetal heart rate tracings and contraction patterns, including the following:

- variable decelerations
- variations in FHR variability
- sinusoidal pattern
- tachysystole
- reactive tracing

Blood pressure cables

Part number	Description
2198879	BCI International TK-1 (6M)
2198879	Criticare Systems Inc. (1100) TK-1 (6M)
2198879	Critikon (Dinamap Plus) TK-1 (6M)
2198887	Datascope DS-1 (6F)
2200955	Datex (AS/3, CS/3, Compact, Cardio Cap II, Critical Care, Light) DX-1 (10F)
2199387	Fakuda Denshi (DS3300 series) FD-2 (12M)
2199682	GE Marquette Medical Corometrics (115, 116, 142, 145, 556) CM-3 (Nicolet round – 12M)
2198893	GE Marquette Medical (PPG/E for M DR) EM-1 (6F)
2198978	GE Marquette Medical (7000 and TRAMAR series only) MQ-2 (8M round)
2199627	GE Marquette Medical (Dash, Eagle, Solar, Tram, and MacLab) MQ-3 (rectangular – 11M)
2198902	Hewlett Packard/Philips (78-300, 78- 500, 78-800, Merlin/Viridia/ Omnicare (HP/Philips M1006B iBP module has a sensitivity of 5 uV/V/ mmHg only. The HP-3 cable should be selected for this application.) HP-3 (12M 5 μ V)
2198916	Hewlett Packard/Philips (78-300, 78- 500, 78-800, Merlin/Viridia/Omnicare) HP-4 (12M 40 μ V)
2198879	Invivo Research TK-1 (6M)
2198879	Ivy Biomedical (400 and 700 series) TK-1 (6M)
2198940	Medical Data Electronics (Escort series) PC-1 (6M)
2198879	North American Drager (Vitalert 2000) TK-1 (6M)
2198940	Physio Control (VSM series) PC-1(6M)
2198879	Protocol System (Propaq series) TK-1 (6M)
2190955	Puritan Bennett PB 240 DX-1 (10F)
2198925	Siemens (SIRECUST series) [SM- 1 and Siemens Medical Transducer Adapter (3368-383-E530U) used to run a single invasive BP channel on the Siemens Medical SC6000 and SC9000 series monitors] SM-1 (10M)
2198879	SpaceLabs (1050, 1700, PCMS series) (SpaceLabs adapters 700-0028-00 and 0120-0551-00 with TK-1 used when testing the new UltraView Command Module) TK-1 (6M)
2392173	Universal unterminated UU-1 (5-Pin DIN one end only)
2198893	Witt Biomedical EM-1 (6F)
5226108	PS8IUP-8001, Corometrics IUP Simulation Cable
5226113	PS8IUP-8000,HP/AG/PH, 50 & 8040 SER, IUP SIMU CABLE

Temperature cables

Part number	Description
2199019	UT-2 standard 1/4 in phone plug (compatible with YSI 700 series – 3 conductor)
2523334	UT-4 Low profile, 1/4 in phone plug, YSI 400 series compatible, two conductor
2199257	HPT-2 temperature adapter (Hewlett Packard) (2 pin, used with UT-1 for HP monitors)

Cardiac output bath/injectate adapters

Part number	Description
2392199	CI-3 cable assembly
2392158	General purpose connector
2199240	COA-1 Cardiac output adapter (Hewlett Packard) (HPT-2 also required for cardiac output simulation on HP patient-monitoring systems)
2199257	HPT-2 Temperature adapter (Hewlett Packard) (2 pin) (COA-1 also required for cardiac output simulation on HP patient-monitoring systems)
4022300	DIN Cardiac Output MARQ EAGLE

About Fluke Biomedical

Fluke Biomedical is the world's leading manufacturer of quality biomedical test and simulation products. In addition, Fluke Biomedical provides the latest medical imaging and oncology quality-assurance solutions for regulatory compliance. Highly credentialed and equipped with a NVLAP Lab Code 200566-0 accredited laboratory, Fluke Biomedical also offers the best in quality and customer service for all your equipment calibration needs.

Today, biomedical personnel must meet the increasing regulatory pressures, higher quality standards, and rapid technological growth, while performing their work faster and more efficiently than ever. Fluke Biomedical provides a diverse range of software and hardware tools to meet today's challenges.

Fluke Biomedical regulatory commitment

As a medical test device manufacturer, we recognize and follow certain quality standards and certifications when developing our products. We are ISO 9001 and ISO 13485 medical device certified and our products are:

- CE Certified, where required
- NIST Traceable and Calibrated
- UL, CSA, ETL Certified, where required
- NRC Compliant, where required

Fluke Biomedical

*We empower our everyday heroes
to focus only on protecting lives.*

Fluke Biomedical
6920 Seaway Boulevard
Everett, WA 98203

For more information, contact us at:
(800) 850-4608 or Fax (440) 349-2307
sales@flukebiomedical.com
flukebiomedical.com

©2011-2013, 2025 Fluke Biomedical.
Specifications subject to change without notice.
6/2025 3983506m-en

Modification of this document is not permitted
without written permission from Fluke Corporation.



LANDAUER®

 **RaySafe™**

