

Manual Supplement

Manual Title: ProSim 4 Users
Print Date: January 2011
Revision/Date: 3, 2/15

Supplement Issue: **2**
Issue Date: 10/19
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 217, 474

On page 4, replace the AC/DC Power Supply with:


AC/DC Power Supply	4766589
--------------------	---------

On page 27, under **General Specifications**, replace the Battery Charger

Battery Charger 110 Vac to 220 Vac, 50/60 Hz input, 6 V/3 A output. For best performance, the battery charger should be connected to a properly grounded ac receptacle.

Change #2, 782

On page 3, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 28, under **Electromagnetic Compatibility (EMC)**, add:

Korea (KCC) Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.