

# Manual Supplement

Manual Title:	ProSim 8 Users	Supplement Issue:	<b>4</b>
Print Date:	January 2011	Issue Date:	7/20
Revision/Date:	3, 3/16	Page Count:	2

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

**Change #1, 631**

On pages 16 & 17 replace Table 10 with:

**Table 10. Arrhythmias by Wave Group**

<b>Arrhythmia</b>	<b>Wave Group</b>
Asynchronous 75 BPM	TV Paced
Asystole	ACLS, Ventricular
Atr-Vent Sequential	TV Paced
Atrial Fibrillation 1 (increased randomness)	Supraventricular, ACLS
Atrial Fibrillation 2 (original)	Supraventricular, ACLS
Atrial Flutter	Supraventricular, ACLS
Atrial PAC	Premature
Atrial Tach	Supraventricular
Atrial 80 BPM	TV Paced
Bigeminy	Ventricular
Sinus Bradycardia	ACLS
Demand Freq Sinus	TV Paced
Demand Occ Sinus	TV Paced
Freq Multi-focal PVCs	Ventricular
Lt Bndl Branch Block	Conduction
Lt Bndl Branch Block	ACLS
Rt Bndl Branch Block	ACLS
Missed Beat at 80 BPM	Supraventricular
Missed Beat at 120 BPM	Supraventricular
Mono V Tach	Ventricular
Mono V Tach (Unstable)	ACLS
Multi-focal PVCs	Premature
Narrow QRS Tach	ACLS
Nodal PNC	Premature
Nodal Rhythm	Supraventricular
Non-Capture	TV Paced
Non-Function	TV Paced
Paroxysmal ATach	Supraventricular
Poly V Tach	Ventricular
Poly V Tach (unstable)	ACLS
PVCs	Ventricular

PVC1 Left Vent	Premature
PVC1 LV Early	Premature
PVC1 LV R on T	Premature
PVC2 Right Vent	Premature
PVC2 RV Early	Premature
PVC2 RV R on T	Premature
Rt Bndl Branch Block	Conduction
Run of PVCs	Ventricular
Sinus Arrhythmia	Supraventricular
Supra V Tach	Supraventricular
Torsade de Pointes	ACLS
Trigeminy	Ventricular
Ventricular Fibrillation 1 (original)	Ventricular, ACLS
Ventricular Fibrillation 2 (faster)	Ventricular, ACLS
Wide QRS Tach	ACLS
1st Deg AV Block	Conduction
2nd Deg AV Block T1	Conduction, ACLS
2nd Deg AV Block T2	Conduction, ACLS
3rd Deg AV Block	Conduction, ACLS

## Change #2, 700

On page A-8, under **Swan-Ganz Procedure**, replace the first sentence with:

After insertion of a Swan-Ganz catheter into a vein (usually the basilic vein of the forearm), the catheter is gently guided by the flow of the blood into the pulmonary artery.

## Change #3, J253

On page 89, under **Non-Invasive Blood Pressure**, in the **NIBP Simulations** section replace the Repeatability with:

Repeatability .....±1 mmHg (Calculated as 1 standard deviation of repeated measurements at maximum pulse size, independent of device under test.)

## Change #4, CK 7/2/2020

On page 4 of the ProSim 8 User Manual (PDF page 16), under **Accessories**

Change the AC/DC Power Supply part number to 4766550.