

Cardiac output simulation for specific makes of monitor

Application Note


By Dennis McMahon, Dennis McMahon Consulting

Theory

The measurement of the volume of blood pumped by the heart is a valuable diagnostic tool in the management of patients undergoing major cardiovascular surgery and for critically ill patients elsewhere in the medical center. Cardiac output (CO) is an indicator of the efficiency of the heart and the status of the entire vascular system. Although methods for determining continuous CO are available, the thermodilution technique is widely used in the Operating Room, the Intensive Care Unit, and the Cardiac Catheterization Lab since it was introduced in the 1970's.

This method requires the placement of a Swan-Ganz catheter in the pulmonary artery. As the name implies, the technique involves briefly decreasing the temperature of the blood in the pulmonary artery by injecting a known quantity of saline into the bloodstream at the right atrium, then detecting the decrease in temperature as the injectate-diluted blood passes a temperature sensor 10 cm downstream at the end of the catheter:


Each injection yields in a time-temperature curve whose area represents the cardiac output:



Note that the vertical axis represents temperature **change**, not actual temperature. The equation for determining cardiac output requires a number of fixed values: the density of blood, the density of injectate, the specific heat of blood, the specific heat of injectate, the volume of injectate, and a correction factor specific to the catheter. The equation is pre-programmed into the monitor in use:

$$CO = K \times \frac{T_i - T_b}{\int \Delta T dt}$$

Where **K** represents the fixed-factor part of the equation, and $\int \Delta T dt$ is the area under the curve.



Corrections are also made to compensate for the extended 'tail' created by the gradual dilution by the injectate (cross-hatched in the graphic above). Manufacturers provide the "computational constant" for each model of pulmonary artery catheter. Depending on the injectate volume and temperature, it may range from 0.132 to 0.608. Ultimately the only variables necessary to input to the monitor are the temperature of the blood (T_b) and that of the injectate (T_i). "Shooting" cardiac outputs is both art and science: the injection must be done quickly, and in one smooth action, in order to create nearly ideal time-temperature curves.

Originally, iced injectate was used in order to assure a large difference between T_i and T_b , to provide better accuracy. With improvements in products and technique, "room temperature" injectate (~ 20 °C) is now more commonly used, unless the ambient room temperature is unusually high.

Simulation procedure for Philips monitors

The thermodilution method is used with any of these Philips modules:


- M1012A, standard or option #C10
- M3012A Hemodynamic Measurement Server Extension, option # C05
- M3014A Capnography Measurement Server Extension, option # C05

Connections

- Connect a Philips M1642A (or M1643A) cable to the CO module or server
- Connect the 4-pin female blood temperature connector from the Philips cable to the 3-pin male on the Fluke Biomedical CI-3 module (Fluke Biomedical # 2392199)
- Connect a 4-pin female injectate temperature adapter (Fluke Biomedical # 2199257) between the Philips cable to the 4-pin male on the CI-3
- Connect the CI-3 to the “Cardiac Output” port of the Prosim 8



Philips M1642A (or M1643A) Cardiac output cable for the CMS or Intellivue series.



Fluke Biomedical CI-3 module. Smaller 3-pin connector is for the blood temperature. Larger (4-pin) connector is for the baseline (injectate) temperature. Dial potentiometer adjusts the injectate temp.



Fluke Biomedical #2199257 Injectate temperature adapter.

Procedure

- On the Philips monitor, enter the C.O. setup window
- On some Philips models, select “Method” and choose “Right Heart”
- Enter a “Computational Constant” of 0.595.
- If necessary, select a “Catheter Size” of 7 F
- If necessary, select an “Injectate Volume” of 10 cc
- On the CI-3 module, adjust the pot for a “Blood Temp” of 24 °C on the Philips screen. (This adjustment may be very critical.)
- On the Prosim 8, press “Special Functions”
- Navigate to “Cardiac Output”
- Navigate to the “Wave” parameter, and select 5.0 L/min
- Navigate to the “Baseline Temp” parameter, and select 37 °C
- Navigate to the “Injectate Temp” parameter, and select 24 °C
- With the “Ready for new measurement” message on the Philips monitor, press the “Start” button on the module (or use the pop-up key “Start CO” on-screen). There will be a tone and the message “Inject now!”
- Immediately press the “Start” softkey on the Prosim 8.

A timer on the Prosim 8 will begin counting-down and a simulated CO waveform will appear on the Philips screen. Within 30 seconds, the Philips will determine and display a CO value, which should be 5.0 L/min \pm 5 %.

After the message “Wait before starting new measurement” changes to “Ready for new measurement”, run at least 2 more measurements. To check a wide range of values, we suggest you also run a CO determination for “Wave” settings of 2.5 L/min and for 10 L/min on the Prosim 8.

If you run CO determinations for iced (0 °C) injectate, don’t forget to change the Computational Constant on the Philips settings to 0.542. (The Prosim 8 will change to this value automatically when you select a 0 °C injectate temp.)



Setup for Philips CMS monitor.



Setup for Philips Intellivue monitor.

Simulation procedure for GE monitors

The thermodilution method is used with any of these GE monitors:

- Dash- series (3000, 4000, & 5000)
- Carescape Patient Data Module

Connections

- Connect Fluke CO adapter cable (Fluke Biomedical # 4022300) from the monitor to the
- Cardiac Output port of the Prosim 8



Fluke Biomedical #4022300 Cardiac output adapter cable for GE.

Procedure

- On the GE monitor, navigate to the CO setup page
- In the "Catheter" box, select Baxter
- In the "Inject Temperature" box, select In-Line
- In the "Size" box, select 7
- In the "Inject Volume" box, select 10
- In the "Computational Constant" box, the CC should default to 0.595
- On the Prosim 8, press "Special Functions"
- Navigate to "Cardiac Output"
- Navigate to the "Wave" parameter, and select 5.0 L/min
- Navigate to the "Baseline Temp" parameter, and select 37 °C
- Navigate to the "Injectate Temp" parameter, and select 24 °C
- Set the rocker switch on the adapter cable to 24 °C
- With "Inject when ready" displayed on the GE monitor, press the "Start" softkey on the Prosim 8

A timer will begin counting down on the Prosim 8 and a simulated CO waveform will appear on the GE monitor screen, and "Computing CO" will display. Within 30 seconds, the monitor will determine and display a CO value, which should be 5.0 L/min \pm 5 %. "CO complete" will display.

After the message "Inject when ready" returns, run at least 2 more measurements. To check a wide range of values, we suggest you also run a CO determination for "Wave" settings of 2.5 L/min and for 10 L/min on the Prosim 8.



Setup for GE monitor.

References

Bowdle, et al: "Cardiac Output" in Biophysical Measurement Series
Spacelabs Medical, 1993

ISBN 0-9627499-2-1

Chan, A: Biomedical Device Technology: Principles and Design
C. Thomas. 2008

ISBN 978-0-398-07700-6

About Fluke Biomedical

Fluke Biomedical is the world's leading manufacturer of quality biomedical test and simulation products. In addition, Fluke Biomedical provides the latest medical imaging and oncology quality-assurance solutions for regulatory compliance. Highly credentialed and equipped with a NVLAP Lab Code 200566-0 accredited laboratory, Fluke Biomedical also offers the best in quality and customer service for all your equipment calibration needs.

Today, biomedical personnel must meet the increasing regulatory pressures, higher quality standards, and rapid technological growth, while performing their work faster and more efficiently than ever. Fluke Biomedical provides a diverse range of software and hardware tools to meet today's challenges.

Fluke Biomedical Regulatory Commitment

As a medical test device manufacturer, we recognize and follow certain quality standards and certifications when developing our products. We are ISO 9001 and ISO 13485 medical device certified and our products are:

- CE Certified, where required
- NIST Traceable and Calibrated
- UL, CSA, ETL Certified, where required
- NRC Compliant, where required

Fluke Biomedical.

Better products. More choices. One company.

Fluke Biomedical

6045 Cochran Road
Cleveland, OH 44139-3303 U.S.A.

Fluke Biomedical Europe

Science Park Eindhoven 5110
5692EC Son, The Netherlands

For more information, contact us:

In the U.S.A. (800) 850-4608 or
Fax (440) 349-2307
In Europe/M-East/Africa +31 40 267 5435 or
Fax +31 40 267 5436
From other countries +1 (440) 248-9300 or
Fax +1 (440) 349-2307
Email: sales@flukebiomedical.com
Web access: www.flukebiomedical.com

©2012 Fluke Biomedical. Specifications subject
to change without notice. Printed in U.S.A.
5/2012 4208004A_EN

**Modification of this document is not permitted
without written permission from Fluke Corporation.**