



Certificate of Compliance

Certificate: 2516806 (LR 44340C)

Master Contract: 162029

Project: 2516806

Date Issued: April 26, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: Gary L. Allen

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, AScT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Saturated oxygen (SpO2) function tester, Model SPOT Light, battery operated rated 3.7V, 10.75Wh.

Notes:

- The above model is Equipment Class II, Pollution Degree 2.
- Provided with Certified AC Adapter rated 100-240 V in, 50/60 Hz, 1.6 Amax, 6 VDC out, 3.5 Amax, 21 VA.
- Provided with Lithium-ion rechargeable battery pack Model BPPS4.

CONDITIONS OF ACCEPTABILITY

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used (for use with AC Adapter).



CSA INTERNATIONAL

Certificate: 2516806 (LR 44340C)

Master Contract: 162029

Project: 2516806

Date Issued: April 26, 2012

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements