

Certificate of Compliance

Certificate: 2595076 Master Contract: 257704

Project: 2595076 **Date Issued:** January 21, 2013

Issued to: Fluke Corporation

6920 Seaway Blvd. Everett, WA 98203

USA

Attention: Mr. Joseph E. Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by:

Lori Sperduti

PRODUCTS

8721 06 - LABORATORY EQUIPMENT 8721-86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Mobile Gas Flow and Pressure Analyzer, Model FLUKE VT305.

For details related to rating, size, configuration, etc. reference should be made to the CSA Certification Record or the descriptive report.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 Safety Requirements for Electrical Equipment for Measurement, Control,

and Laboratory Use - Part 1: General Requirements (Third Edition)

ANSI/UL Standard 61010-1 Electrical Equipment For Measurement, Control, and Laboratory Use;

Part 1: General Requirements (Third Edition)

ANSI/ISA-61010-1 (82.02.01) Electrical Equipment For Measurement, Control, and Laboratory Use;

Part 1: General Requirements (Third Edition)

DQD 507 Rev. 2009-09-01 Page 1

Certificate: 2595076 **Project:** 2595076



Master Contract: 257704 **Date Issued:** January 21, 2013

MARKINGS

Submittor's name and/or CSA file number 257704, model designation and the CSA Mark appear in a permanent manner on each unit.

DQD 507 Rev. 2009-09-01 Page 2



Supplement to Certificate of Compliance

Certificate: 2595076 Master Contract:

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description	
2595076	January 21, 2013	Original Certification.	
		Submittor	Listee
		IMTMedical	Fluke Corporation
		MC 220541	MC 257704
		CITREX Hx	FLUKE VT305

DQD 507 Rev. 2009-09-01 Page 1