Report # EMC2010May20-1303

# ELECTROMAGNETIC COMPATIBILITY TEST REPORT for the

# Fluke Biomedical VTMobile Gas Flow Analyzer



The Fluke Biomedical VTMobile was tested to the following standards at the EMC laboratories of Fluke Corporation. 6920 Seaway Blvd Everett WA 98203

## EN 61326-1:2006, EN61326-2-1:2006, CISPR 11:2004 Class A Emissions and Immunity

The Fluke Biomedical VTMobile passes test requirements for equipment used for:

	Industrial Locations	Industrial Locations				
	Non-Domestic Us	e (Class A)	□ Domestic Use (Class B)			
	Class A equipment is equipment suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.					
Prepared by: Mary Schurch Gary Allen Test Engineer			Date: <u>19-N</u>	<u>May-2010</u>		
ŀ	Approved by: A					

Thomas Smith Thomas & South Test Engineer Sr.

Date: Jun 17, 2010



### III. Test Results

The Fluke Biomedical VTMobile Gas Flow Analyzer was tested to the following Electromagnetic Compatibility [EMC] requirements:

#### Adapted from CISPR 11 Table 2a Emissions Limits for Class A Equipment

	Port	Frequency MHz	Limits	Standard	Pass/Fail
	Enclosure 30 to 230		40dB (uV/m) quasi peak, measured at 10 meters.		Pass
	Eliciosule	230 to 1000	47dB (uV/m) quasi peak, measured at 10 meters.		Pass
		0.15 to 0.5	79dB (uV/m) quasi peak, 66dB (uV/m) average.	CISPR 11	Pass
	AC mains	0.5 to 5.0	73dB (uV/m) quasi peak, 60dB (uV/m) average.		Pass
		5 to 30	73dB (uV/m) quasi peak, 60dB (uV/m) average.		Pass

#### Adapted from EN 61326-1:2006 Table 1 Basic immunity test requirements

Port	Phenomenon	Basic standard	Test value	Minimum Criteria	Pass/Fail
	ESD	EN 61000-4-2	4 kV/4 kV contact/air	В	Pass
	EM Field	EN 61000-4-3	3 V/m (80 MHz to 1 GHz)	А	Pass
Enclosure	EM Field	EN 61000-4-3	3 V/m (1,4 GHz to 2 GHz)	А	Pass
	EM Field	EN 61000-4-3	1 V/m (2,0 GHz to 2,7 GHz)	А	Pass
	Voltage dip	EN 61000-4-11	0 % during half cycle	В	Pass
	Voltage dip	EN 61000-4-11	0 % during 1 cycle	В	Pass
AC power	Voltage dip	EN 61000-4-11	70 % during 25/30 <sup>e)</sup> cycles	С	Pass
(including protective	Short interruptions	EN 61000-4-11	0 % during 250/300 <sup>e)</sup> cycles	С	Pass
earth)	Burst	EN 61000-4-4	1 kV (5/50 ns, 5 kHz)	В	Pass
	Surge	EN 61000-4-5	0,5 kV <sup>a)</sup> /1 kV <sup>b)</sup>	В	Pass
	Conducted RF	EN 61000-4-6	3 V (150 kHz to 80 MHz)	А	Pass
	Burst	EN 61000-4-4	1 kV(5/50 ns, 5 kHz)	В	Pass
DC power <sup>a)</sup>	Surge	EN 61000-4-5	0,5 kV <sup>a)</sup> / 1 kV <sup>b)</sup>	В	Pass
(including protective earth)	Conducted RF	EN 61000-4-6	3 V (150 kHz to 80 MHz)	А	Pass
I/O signal/ control	Burst	EN 61000-4-4	0,5 kV <sup>a)</sup> (5/50 ns, 5 kHz)	В	Pass
(including lines	Surge	EN 61000-4-5	1 kV <sup>b), c)</sup>	В	Pass
connected to functional earth port)	Conducted RF	EN 61000-4-6	3 V <sup>d)</sup> (150 kHz to 80 MHz)	А	Pass
I/O cignal/	Burst	EN 61000-4-4	1 kV(5/50 ns, 5 kHz)	В	Pass
control	Surge	EN 61000-4-5	0,5 kV <sup>a)</sup> / 1 kV <sup>b)</sup>	В	Pass
connected directly to mains supply	Conducted RF	EN 61000-4-6	3 V (150 kHz to 80 MHz)	А	Pass

a) Line to line.

b) Line to earth (ground).

c) Only in the case of long-distance lines (see 3.6).

d) Only in the case of lines > 3 m.

e) "25/30 cycles" means "25 cycles for 50 Hz test" and "30 cycles for 60 Hz test".

EMC TEST REPORT	FLUKE CORPORATION	
Fluke Biomedical VTMobile Gas Flow Analyzer	28-Sep-2012	PAGE v
COMPANY CONFIDENTIAL COMPANY CONFIDENTIAL COMPANY CONFIDENTIAL THIS TEST DEDODT IS COMPANY CONFIDENTIAL ANY DEDDODI/CTION. IN WHOLE OD IN DADT. IS PROHIBITED WITHOUT	PRICE WRITTEN APPROVAL FROM FLUKE CORPORATION. COMPANY CON	EIDENTIAL COMPANY CONFIDENTIAL COMPANY CONFIDENTIAL