



# Certificate of Compliance

**Certificate:** 2578554

**Master Contract:** 162029

**Project:** 2578554

**Date Issued:** November 22, 2012

**Issued to:** Fluke Corporation  
6920 Seaway Blvd  
Everett, WA 98203  
USA

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Cindy Johnson*

**Issued by:** Cindy Johnson, ASCT.

## **PRODUCTS**

**CLASS 3631 06** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 86** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Model VT Plus (Gas Flow Analyzer), cord connected, rated 100-240 V ac, 50/60 Hz, < 132 VA.

Model BP Pump 2 (Non-Invasive Blood Pressure Monitor Analyzer) , cord connected, rated 100-240 V ac, 50/60 Hz, < 60VA.

## **Notes:**

The above models are Equipment Class I, Pollution Degree 2, Measurement Category none.



**Certificate:** 2578554

**Master Contract:** 162029

**Project:** 2578554

**Date Issued:** November 22, 2012

---

### **APPLICABLE REQUIREMENTS**

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Reference Standards:

IEC/EN 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

### **CONDITIONS OF ACCEPTABILITY**

1. The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.