

Manual Supplement

Manual Title: BP Pump 2 Operator Supplement Issue: 4
Part Number: 2196592 Issue Date: 6/17
Print Date: June 2007 Page Count: 2
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 55935

On page 1-11, following **Pressure Generation** add:

Dynamic NIBP Simulation Repeatability..... ± 2 mmHg (at maximal pulse size independent of device under test)

Change #2, 56150, 56170



On page 1-11, under **Pressure Measurement**, change the High-Accuracy Version to the following:

High-Accuracy Version..... ± 0.7 mmHg (0.09 kPa)

Change #3, 63302, 65378, 422, 216

On page 1-4, delete the entire **General Safety Considerations** section.

On page 1-4, in the **Symbols** Table delete the **ETL** Symbol and add:

	Conforms to relevant North American Safety Standards.
	Conforms to relevant South Korean EMC Standards.

On page 1-10, following **Specifications**, delete **Mains Voltage**, **Environmental Conditions** and add the new section:

General Specifications

Power Line100 - 240 V ac 50/60 Hz, 60 VA
 Operating Temperature15 °C to 40 °C
 Storage Temperature-20 °C to +65 °C
 Relative Humidity90 % max
 Operating Altitude.....2000 m
 SafetyIEC 61010-1: Overvoltage CAT II,
 Pollution Degree 2

Electromagnetic Compatibility (EMC)

International.....IEC 61326-1: Basic, IEC 61326-2-2
 CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for *domestic purposes*. *There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.*

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object. The equipment may not meet the immunity requirements of this standard when test leads and/or test probes are connected.

Korea (KCC)..... Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC)..... 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.

CE Mark..... BP Pump 2L only.